

## How to adjust MD-3025/ MD-3025-IVS focus?

1. Unscrew two screws



2. Pull the cable gently



3. Remove the top cover





4. Pull the robber ring gently.



5. Adjust the focus



**The desiccant replacement**

Please replace the desiccant to avoid the moisture in the MD-3025.

1. There are two new desiccants in the box.



2. Unscrew two screws





3. Replace two desiccants.

